



WASHINGTON STATE PATROL
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

South Side

DATE: July 12 2006 INITIALS: JEC OWNER: WSP
DISTRICT: 6 ADDRESS: I-90 West MP 80
LOCATION: Ch Elum No 53 COUNTY: Kittitas
MAKE: DSE TYPE: Truck scale ACCURACY CLASS: M4
MODEL: 651 ID. NO: 202173 CAPACITY: 100,000 SECTIONS: 2
MINIMUM GRADUATIONS: 20 REG. ELEMENT: Digital PLATFORM: Dual 12' X 20'

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
<u>Platform</u>		<u>No 1</u>		<u>Platform</u>		<u>No 2</u>	
<u>Sec 1</u>	<u>20,000</u>	<u>20,000</u>	<u>0</u>	<u>Sec 1</u>	<u>20,000</u>	<u>20,000</u>	<u>0</u>
<u>Center</u>	<u>20,000</u>	<u>20,000</u>	<u>0</u>	<u>Center</u>	<u>20,000</u>	<u>20,000</u>	<u>0</u>
<u>Sec 2</u>	<u>20,000</u>	<u>20,000</u>	<u>0</u>	<u>Sec 2</u>	<u>20,000</u>	<u>20,000</u>	<u>0</u>

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as is shown on the attached "Certificate of Accuracy", No. 7 2441-1 and 2

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David E Cronin Date: 8/3/2006



State of Washington
County of THURSTON
Signed and sworn to (or affirmed) before
me on 8-3-2006 by
Sylvia S. Lewallen
(Signature of notary public)
SYLVIA S. LEWALLEN
Printed Name of Notary
My appointment expires 3-28-2007

3000-362-021 (R) (10/98)

DISTRIBUTION: WHITE - COURT GREEN - DISTRICT CANARY - PROPERTY MANAGEMENT PINK & GOLDENROD - SERVICE TRUCKS